



**Single Channel Electronic pipette LD-
SEP101L**

Labodam Equipment Ltd.
info@labodam.com | www.labodam.com

Single Channel Electronic pipette LD-SEP101L

Features:

- Dual charging modes for longer operation time
- Digital control and multifunctioning Pipette
- Autoclavable (Lower part)
- Adjustable speeds for aspiration and dispensing
- Self-calibration, can be connected to a PC via USB for calibration
- Accuracy Error and Precision Error is given according to ISO8655-2 standard

Specifications:

Increment	0.1 μ L
Test Volume	50 μ L <hr> 5 μ L
Volume Range	5 – 50 μ L
Accuracy Error	\pm 0.80 % <hr> \pm 3.00 %
Precision Error	0.3 % <hr> 2.50 %