



*Fixed Volume Fully Autoclavable  
Pipettes FVP102L*

## Description :

## Specifications :

Test Volume	20 µL
Accuracy error	0.6 %
Precision error	0.5 %

## Features :

Autoclavable at 121 Â°C

Material: Polycarbonate

Excellent durability and chemical resistance

Ergonomic design provides excellent operating experience

Easy-to-read volume display

Trouble-free calibration and maintenance



**Labodam Equipment Ltd**

18a Melton Road Leicester LE4 5EA  
United Kingdom

[www.labodam.com](http://www.labodam.com) // [info@labodam.com](mailto:info@labodam.com)