



*Fixed Volume Fully Autoclavable
Pipettes FVP109L*

www.labodam.com // info@labodam.com

Description :

Specifications :

Test Volume	1000 μ L
Accuracy error	0.3 %
Precision error	0.2 %

Features :

Autoclavable at 121 $\text{Å}^{\circ}\text{C}$

Material: Polycarbonate

Excellent durability and chemical resistance

Ergonomic design provides excellent operating experience

Easy-to-read volume display

Trouble-free calibration and maintenance

The logo for Labodam features a stylized green and white icon of a laboratory flask or beaker with three small dots above it, followed by the word "labodam" in a bold, black, sans-serif font.

Labodam Equipment Ltd

18a Melton Road Leicester LE4 5EA

United Kingdom

www.labodam.com // info@labodam.com