



Single Channel Electronic pipette
SEP103L

Description :

Specifications :

Volume Range	100 - 1000 μL
Increment	5 μL
Test Volume	1000 μL
	100 μL
Accuracy Error	$\pm 0.60\%$
	$\pm 3.00\%$
Precision Error	0.20 %
	0.60 %

Features :

Dual charging modes for longer operation time

Digital control and multifunctioning Pipette

Autoclavable (Lower part)

Adjustable speeds for aspiration and dispensing

Self-calibration, can be connected to a PC via USB for calibration

Accuracy Error and Precision Error is given according to ISO8655-2 standard



Labodam Equipment Ltd

18a Melton Road Leicester LE4 5EA

United Kingdom

www.labodam.com // info@labodam.com