



*Variable Volume Fully Autoclavable
Pipettes VVP103L*

Description :

Specifications :

Volume Range	5 - 50 μL
Increment	0.5 μL
Test Volume	50 μL
	25 μL
	5 μL
Accuracy error	0.60 %
	0.90 %
	2.00 %
Precision error	0.30 %
	0.60 %
	2.00 %

Features :

Autoclavable at 121 $^{\circ}\text{C}$

Material: Polycarbonate

Available in 0.1 μL to 5000 μL volume

Ergonomic design provides excellent operating experience

Easy-to-read volume display

Trouble-free calibration and maintenance



**18a Melton Road Leicester LE4 5EA
United Kingdom
www.labodam.com // info@labodam.com**