



## *Variable Volume Fully Autoclavable Pipettes LD-VVP107L*

## Description :

## Specifications :

Volume Range	100 - 1000µL
Increment	5 µL
Test Volume	1000 µL
	500 µL
	100 µL
Accuracy error	0.60 %
	0.70 %
	2.00 %
Precision error	0.20 %
	0.25 %
	0.70 %

## Features :

Autoclavable at 121 °C

Material: Polycarbonate

Available in 0.1 µL to 5000 µL volume

Ergonomic design provides excellent operating experience

Easy-to-read volume display

Trouble-free calibration and maintenance