



Variable Volume Pipette VVP201L

Description :

Specifications :

Volume Range	0.5 - 10 μL
Increment	0.1 μL
Test Volume	10 μL
	5 μL
	1 μL
Accuracy error	1.00 %
	1.50 %
	2.50 %
Precision error	0.80 %
	1.50 %
	1.50 %

Features :

Autoclavable at 121 $^{\circ}\text{C}$

Material: Polycarbonate

Available in 0.1 μL to 10000 μL volume

Ergonomic design provides excellent operating experience

Easy-to-read volume display

Trouble-free calibration and maintenance



Labodam Equipment Ltd

**18a Melton Road Leicester LE4 5EA
United Kingdom**

www.labodam.com // info@labodam.com