



*Variable Volume Pipette VVP202L*

## Description :

## Specifications :

Volume Range	2 - 20 $\mu\text{L}$
Increment	0.5 $\mu\text{L}$
Test Volume	20 $\mu\text{L}$
	10 $\mu\text{L}$
	2 $\mu\text{L}$
Accuracy error	0.90 %
	1.20 %
	3.00 %
Precision error	0.40 %
	1.00 %
	2.00 %

## Features :

Autoclavable at 121  $^{\circ}\text{C}$

Material: Polycarbonate

Available in 0.1  $\mu\text{L}$  to 10000  $\mu\text{L}$  volume

Ergonomic design provides excellent operating experience

Easy-to-read volume display

Trouble-free calibration and maintenance

**Labodam Equipment Ltd**

**18a Melton Road Leicester LE4 5EA  
United Kingdom**

**[www.labodam.com](http://www.labodam.com) // [info@labodam.com](mailto:info@labodam.com)**