



Variable Volume Pipette VVP205L

Description :

Specifications :

Volume Range	20 - 200 µL
Increment	1 µL
Test Volume	200 µL
	100 µL
	20 µL
Accuracy error	0.60 %
	0.80 %
	3.00 %
Precision error	0.15 %
	0.30 %
	1.0 %

Features :

Autoclavable at 121 °C

Material: Polycarbonate

Available in 0.1 µL to 10000 µL volume

Ergonomic design provides excellent operating experience

Easy-to-read volume display

Trouble-free calibration and maintenance



**18a Melton Road Leicester LE4 5EA
United Kingdom
www.labodam.com // info@labodam.com**